


A New Millennium for Inhaler Technology

Anthony J. Hickey* and Craig A. Dunbar



In recent years dramatic developments have occurred in the technology associated with aerosol delivery of drugs to the lungs. Propellant-based metered-dose inhalers have entered a new phase in which chlorofluorocarbons have been joined by so-called environmentally friendly propellants as a means of aerosol propulsion. Active emission dry powder inhalers have been manufactured, and may supersede those based on the passive inspiratory flow of the patient. Self-contained, hand-held, aqueous spray systems are being introduced and are capable of competing with metered-dose inhalers and dry powder inhalers. The current status of pharmaceutical inhalation aerosol technology indicates a bright future for drug delivery to the lungs, not only as the site of action but also as a route of administration.

The development of the first pressurized metered-dose inhaler (pMDI) for asthma therapy (Riker Laboratories, 1956) was a major advance in the use of aerosols for delivery to the lungs (1). Chlorofluorocarbon (CFC) propellants were used to disperse drug particles or droplets from the inhalers in sizes suitable for lung deposition. The metering of nominal drug dose by a novel valve developed by Meshberg was central to the success of this device (2). Currently, the pMDI is the most frequently prescribed inhaler for the treatment of asthma. However, it is under increasing scrutiny because of the postulated contribution of CFCs to ozone depletion and global warming (3). Reformulation using so-called ozone-friendly propellants is the only alternative approach based on retaining pMDI technology. In response to the limited number of CFC propellants and reformulation opportunities, alternative devices to the pMDI have been given greater consideration. A number of research reports and patents describing the performance of dry powder inhalers (DPIs) indicate their leading position as a truly alternative technology. Conventional nebulizers are effective devices in terms of droplet dispersion and delivery to the peripheral regions of the lung, but they are as portable as the other systems described above and are popular for use outside the home or hospital. However, formulators are developing a new generation of hand-held, portable devices that will extend the use of aqueous aerosols as an alternative to pMDIs and DPIs.

In the 40 years since the development of the first pressurized MDIs, the possibility of using the lung as a target route of administration for medicines has increased dramatically. The pages of *Pharmaceutical Technology* have of-

Anthony J. Hickey, PhD, is an associate professor, and **Craig A. Dunbar, PhD**, is a post-doctoral fellow in the Dispersed Systems Laboratory, School of Pharmacy, Beard Hall CB# 7360, University of North Carolina, Chapel Hill, NC 27599, tel. (919) 962-0223, fax (919) 966-6919, e-mail (AHICKEY@UNC.EDU). Dr. Hickey is a member of the Editorial Advisory Board of *Pharmaceutical Technology*.

*To whom all correspondence should be addressed.




Figure 1: Comparison of predicted initial drop sizes issued from a pMDI for various propellants.

flected the state of the art or current debates regarding the means of delivery and characterization of these systems. The following overview addresses the role of various devices in the treatment of respiratory diseases and considers their future contributions to drug delivery to and via the respiratory tract. Methods of particle-size characterization are presented as this has been a topic of some debate in the past decade.

PRESSURIZED METERED-DOSE INHALERS

The pressurized metered-dose inhaler offers the most convenient, versatile, and cost-effective means of aerosol drug delivery available commercially. The device typically discharges more than 200 doses, producing a respirable spray of finely dispersed drug. The primary component of the pMDI is the formulation, the other components being the metering valve, actuator, and container.

The formulation most frequently consists of micronized drug particles suspended in a combination of propellants. Solution formulations are also available, but they represent a small proportion of the pressure-packaged aerosol market. Propellant blends have traditionally been a combination of chlorofluorocarbons (CFC-12, 11, and 114) providing the energy for atomization by virtue of their vapor pressure and leading to the production of respirable droplets. These propellants also have the important pharmaceutical characteristics of low toxicity, chemical stability, low cost, and good organoleptic properties.

Molina and Rowland first suggested in 1974 that the decomposition of CFCs by ultraviolet radiation in the upper atmosphere could lead to a build-up of chlorine and subsequent depletion of stratospheric ozone by the formation of chlorine monoxide (4). This was confirmed in 1985 with the discovery of an ozone hole in the atmosphere above Antarctica (5). The Montreal Protocol initiated international agreement leading to a ban on the production of all CFCs from the beginning of 1995 (6,7). As part of the effort to prevent further ozone depletion, the pharmaceutical industry responded by attempting to rapidly reformulate their pMDI devices with environmentally friendly propellants. It is not yet clear that all pMDI formulations can be

reformulated in alternative propellants. Intermediate solutions to the reformulation problems have been suggested (8,9). Consequently, CFC propellants are expected to be retained for delivery of some compounds for the foreseeable future.

Open discussions of the reformulation issues for pMDIs by propellant manufacturers, the pharmaceutical industry, and academicians were a prelude to an "essential use" exemption recognition of the complexities of reformulation (10,11). The situation will be reviewed in the near future, depending on the successful reformulation of pMDI products and the necessary approval by regulatory authorities. Formulation of albuterol with an alternative propellant (Airomir; 3M Pharmaceuticals, St. Paul, MN), currently marketed in the European Union and United States, may well have a direct impact on the future of CFC propellant-based products.

The alternative propellant candidates that have been assessed for toxicological approval are hydrofluorocarbons (HFCs) 134a and 227. The use of alternative propellants depends upon the ability to disperse droplets or particles in similar or smaller sizes than are generated by the current CFC products. Figure 1 shows the predicted mass median diameters ($D_{0.5}$) produced at the exit of the pMDI actuator for each prospective alternative propellant using an actuator flow model (12). These results are compared with the drop sizes produced by a traditional CFC-11/12 (28:72% w/w) propellant combination delivered via a 63-µm DF60 metering valve (Perfect Valois, Greenwich, CT). The results predict that the alternative propellants produce smaller initial drop sizes than do the CFC-based propellants and therefore can be considered as suitable replacements. Reformulation of pMDIs has required modification of metering valve design and development of new surfactants and cosolvents. As these technical difficulties are addressed, a number of alternative propellant devices will likely be available.

One of the most significant drawbacks of the pMDI has been patient coordination at the correct point in the inspiratory effort. Pediatric and elderly patients are most susceptible to problems of device coordination. This is now being overcome with the introduction of breath-actuated devices such as the Autohaler (3M Pharmaceuticals, St. Paul, MN) shown in Figure 2a. This device incorporates a diaphragm that actuates the pump at a given inspiratory pressure, thus negating the need for hand-inspiration coordination by the patient. More "intelligent" devices are being developed. SmartMist (Aradigm, Hayward, CA), shown in Figure 2b, is a hand-held device that integrates breath-actuation capabilities, a miniature pneumotachograph, and microprocessor so that a drug bolus can be delivered at a preprogrammed point in the inhalation cycle. Such devices certainly have a role to play in future pMDI development.

DRY POWDER INHALERS

Dry powder inhalers are of increasing interest for the delivery of drugs to the lungs. The basis of aerosol delivery by these systems is the combination of the powder product, the metering system, and the method of dispersion. The first system for delivery of dry powder drugs in modern times was the Spinhaler (Fisons, Rochester, NY) (13). This system has a unique delivery mechanism that combines rotation with vibration to generate the aerosol (14).




Figure 2: Examples of breath-actuated and microprocessor-controlled pMDI devices: (a) Autohaler (3M Pharmaceuticals) and (b) SmartMist (Aradigm).

In general, the powder product may consist of drug alone or blended with excipient (e.g., lactose). The excipient acts as a carrier with a size range of 30–60 μm , which is much larger than that required (<5 μm) for delivery to the lung (15). These carrier particles aid in the dispersion of the drug. The ease of

dispersion of drug particles relates to their particle size distribution, morphology, and surface characteristics. These are related to the forces of interaction between particles (16). A number of studies of basic powder properties and turbulent flow have been performed (17,18). However, further investigation may also be required for dry powder inhalers to fulfill their potential as alternatives to the pMDI (19).

The metering systems used in DPIs include hard gelatin capsule unit doses (Spincaps; Fisons, Rochester, NY, and Rotacaps; Glaxo Wellcome, Research Triangle Park, NC); blister packaged multiple unit doses (Diskhaler; Glaxo Wellcome, Research Triangle Park, NC) and reservoir systems (Turbuhaler; Astra; Lund, Sweden). The stability of the powder in these systems must be considered during the formulation process. Moisture ingress is a major source of both physical and chemical instability (20). It is often necessary to include desiccants to remove water from the immediate environment of the powdered drug.

The method of dispersion of commercially available products is passive inhalation (Spinhaler, Rotahaler, Diskhaler, and Accuhaler; Glaxo Wellcome, Research Triangle Park, NC, and Turbuhaler and Inhalator; Boehringer Ingelheim, Ridgefield, CT). This has the advantage of the drug source being dispersed by patients' own inspiratory effort; its disadvantage is requiring patient's optimum inhalation flow rate to effectively disperse powder (21). Products currently under development use active dispersion mechanisms, for example, Spiros (Dura, San Diego, CA) and the Inhale Deep Pulmonary Drug Delivery System (Inhale, Palo Alto, CA). These devices are shown in Figures 3a and 3b, respectively. They offer the advantage that aerosol generation is independent of the patient's inspiratory flow. Thus, optimal lung deposition and therapeutic effect are achieved.

Pharmaceutical dry powders for lung delivery are commonly manufactured by jet milling and spray drying. Jet milling reduces the size of particles by attrition using high-velocity opposing air jets (22). Spray drying produces a dry powder by controlled evaporation of an atomized solution (23). Each of these manufacturing techniques produces particles with different characteristics that can alter the dispersion properties of bulk powder.

A current area of interest in the production of respirable particles is supercritical fluid manufacture. This is a controlled crystallization technique in which the supercritical fluid can be presented as either a solvent or antisolvent (23–29). The advantage of supercritical fluid manufacture in the production of pharmaceutical powders is the elimination of the high-temperature processes (attrition and heat transfer) that might damage the product (e.g., by denaturation of proteins or other labile molecules). This technique is known to produce particles of uniform size and morphology, which are attractive characteristics for aerosol delivery.

NEBULIZERS

Nebulizers offer the simplest and most effective means of droplet delivery to the peripheral regions of the lung. The principles used in nebulization of aqueous droplets fall into two categories: airblast and ultrasonic aerosol generation. All nebulizers work on the principle that a high-pressure air stream produces a local pressure drop, thus drawing the fluid in.




Figure 3: Examples of active dispersion DPI devices: (a) Spiros (Dura) and (b) Inhale deep pulmonary drug delivery system (Inhale).

airstream by capillary action. The fine liquid stream is then disintegrated by shear forces. Ultrasonic nebulizers work by imposing a rapidly oscillating waveform onto a liquid film via an electromechanical vibrating surface. At a given amplitude the waveform becomes unstable, disintegrating the liquid film and producing small droplets.

An advantage of the nebulizer is the relative ease of formulating aqueous solutions. In the past, the principal drawback of

the nebulizer was its lack of mobility because it required a compressed gas source (airblast) or electrical power unit (ultrasonic). A convenient and portable vibrating aperture nebulizer that uses a small electrical power unit (shown in Figure 4) is currently under development (Aerogen, Santa Clara, CA). This may be the first step toward a new generation of nebulizers.

One exciting prospect that has been in development during the past 5 years is that of a truly portable, hand-held nebulizer that will deliver aqueous solutions complementing the powder and pressure-packaged devices. Two new systems, the BiNeb (Boehringer Ingelheim, Ingelheim am Rhein, Germany) and AERx (Afadigm, Hayward, CA), operate on different principles that allow a single-unit dose of drug to be administered. The BiNeb consists of a nozzle that has two coincident outlet channels ($\sim 8 \mu\text{m}$) through which the liquid dose is accelerated, resulting in impaction of the two streams and atomization. The energy for atomization is provided by a compressed spring and metering pump (30). AERx is based on the principle of a unit-dose disposable blister pack (31). A mechanical actuator forces the liquid dose in the blister container through a multiorifice nozzle, producing an aerosol (32). A heating element is provided to control the relative humidity of the inhaled air, and it includes microprocessor-controlled delivery technology, similar to that of the SmartMist.

CHARACTERIZATION METHODS

Particle-size characterization of respirable aerosols has been the subject of much discussion because of the requirement of *in vitro* testing and the tendency to extrapolate data to lung deposition. Methods for characterizing inhalation aerosols are diverse. They have been reviewed by the Particle Size Subcommittee of the AAPS Aerosol Group and include impaction methods, cascade impaction, laser diffraction, time-of-flight aerosol beam spectrometry, holography, microscopy, right-angle light scattering, and phase-Doppler particle analysis (33-40).

The primary method for aerosol characterization is inertial impaction, which provides information on the aerodynamic particle size and distribution by inertial differentiation of particles. A number of pharmacopeias have adopted this method. The value of inertial impaction lies in its ability to provide information on drug content of the whole aerosol, along with particle size and distribution.

Inertial impactors provide quantitative data on device performance that are extremely useful in setting product specifications but are often interpreted in terms of the potential for lung deposition. After considerable debate, discussion continues on the legitimacy of some interpretations of impaction data. Most of this discussion focuses on selection of cutoff diameters to describe the aerosol distributions, representative sampling flow rates, sampling inlet geometries. One interesting observation is that the human throat exhibits a broad cutoff collection efficiency across a range of particle diameters (41). This is illustrated in Figure 5, which shows the retention efficiency of the throat predicted by a semiempirical model for an average inspired flow rate of 64.5 L/min (42). The two-stage liquid impinger follows a broad cutoff collection efficiency, as shown by the retention curve in Figure 5 for a flow rate of 60 L/min . The retention efficiency curve similar to the one obtained at

mentally has been predicted using computational fluid dynamics (CFD), also shown in Figure 5 (43,44). Thus, under these defined conditions the collection efficiency of the two-stage liquid impinger could be considered similar to the deposition characteristics in the oropharynx. The concept of the head as a blunt sampling inlet is not novel, and it is possible to design impactors to approximate a predetermined deposition curve (45,46). Nevertheless, anyone interested in the lung deposition of aerosols is referred to direct studies of this phenomenon. Radioisotopes may be used to quantify lung deposition, and predictive therapeutic models have been developed (47,48).

It is essential for thorough characterization that methods must not oversimplify the description of the aerosol. The complex interactions occurring within an aerosol may require a knowledge of the particle-size distribution; location of the drug; velocity of the emitted particles or droplets; fluid flow, heat, and mass transfer rates; and plume dimensions. The uncoupling of these interactions can lead to exhaustive experimental designs, and for this reason there may well be a role for computational modeling of the aerosol to provide an insight into areas that are difficult to delineate experimentally.

Standard methods of particle-size analysis for pMDIs and DPIs are being developed (49-51). In contrast, standard testing procedures have not been specified to evaluate nebulizer performance. The nebulizer has a number of components (e.g., reservoir, outlet, tubing, and mouthpiece/face mask); therefore, the definition of a benchmark nebulizer system is required because modifications in any of these parameters can significantly alter the estimated output characteristics (52). A system of this nature should be accompanied by definitions of standard operating conditions including airflow rates, adaptors, sampling tubes, and sampling times.

CONCLUSION

Many exciting developments have taken place in the technology associated with the delivery of medicinal aerosols. The Montreal Protocol brought about an irreversible change in the approach to research and development of inhalers, directly for the pMDI and indirectly for both the DPI and nebulizers (6). Multiple issues are now being addressed to improve these delivery systems in areas that were considered satisfactory before the introduction of the Montreal Protocol. Investigations into the basic mechanisms of pMDI spray production, the development of new technologies, and approaches to device development are now taking the propellant-driven inhaler into the 21st century (53-55). The same is true of the DPI, as the mechanisms of powder dispersion are better understood (14,16), along with the development of more exotic formulations and novel devices. Nebulizers will have a role to play in the future as technology transforms the device into a compact, self-contained aerosol delivery device.

New devices may be designed to satisfy patients' needs. Feedback into the design process with regard to the practical and esthetic aspects of the inhalers may become commonplace, thus making them simple to use, and capable of monitoring and modifying therapy by collating information on usage (e.g., counters, pulmonary function testing, etc.) (56). Aerosol formulations containing proteins, peptides, and nucleic acids for lo-




Figure 4: Aerogen portable nebulizer (Aerogen).




Figure 5: Retention efficiency curves for the two-stage liquid impinger and predicted oropharyngeal deposition (42-44).

cal and systemic therapy are already under development in parallel with novel devices for their delivery (57). However, significant challenges need to be overcome in this area, including formulation development to establish a general understanding of the complex interactions of proteins and peptides with propellants and additives, as well as degradation during manufacture, storage, and delivery (58).

Pharmaceutical Technology has always been an open forum for presentation of current and novel concepts in aerosol delivery. It will continue to be a useful medium through which to

encourage discussion and communicate information to the pharmaceutical community on new developments in the field of inhalation aerosol technology.

REFERENCES

1. C.G. Thiel, "From Susie's Question to CFC-Free: An Inventor's Perspective on 40 Years of MDI Development and Regulation," in *Respiratory Drug Delivery V*, P.R. Byron, R.N. Dalby, and S.J. Farr, Eds. (Interpharm Press, Buffalo Grove, 1996), pp. 115-123.
2. P. Meshburg, "Aerosol Containers and Valves Thereof," US Patent No. 2721010, October 1955.
3. P.S. Zurek, "Producers, Users, Grapple with Realities of CFC Phaseout," *Chem. Eng. News* 67 (30), 7 (1989).
4. M.J. Molina and F.S. Rowland, "Stratospheric Sink for Chlorofluoromethanes: Chlorine Atom-Catalyzed Destruction of Ozone," *Nature* 249, 810-812 (1974).
5. J.C. Farman, B.G. Gardiner, and J.D. Shanklin, "Large Losses of Total Ozone in Antarctica Reveal Seasonal ClO₂/NO₂ Interaction," *Nature* 315, 207-210, (1985).
6. Montreal Protocol on Substances That Deplete the Ozone Layer, 26 ILM 1541 (1987).
7. Council Regulation (EEC), Council Regulation on Substances that Deplete the Ozone Layer, No. 594/91 of 4 March 1991 and amended by No. 3952/92 of 30 December 1992.
8. R.N. Dalby et al., "CFC Propellant Substitution: P-134a as a Potential Replacement for P-12 in MDIs," *Pharm. Technol.* 14 (3), 26-33 (1990).
9. R.N. Dalby and P.R. Byron, "Metered-Dose Inhalers Containing Flammable Propellants: Perspectives and Some Safety Evaluation Procedures," *Pharm. Technol.* 15 (10), 34-66 (1991).
10. F.X. Fischer et al., "CFC Propellant Substitution: International Perspectives," *Pharm. Technol.* 13 (9), 44-52 (1989).
11. Montreal Protocol, Report of the Technology and Economic Assessment Panel Including Recommendations on Nominations for Essential Use Production/Consumption Exemptions for Ozone Depleting Substances, March 1994.
12. C.A. Dunbar, A.P. Watkins, and J.F. Miller, "A Theoretical Investigation of the Spray Issued from a pMDI," *Atomization and Sprays*, in press, 1997.
13. J.H. Bell, P.S. Hartley, and J.S.G. Cox, "Dry Powder Aerosols I: A New Powder Inhalation Device," *J. Pharm. Sci.* 60 (10), 1539-1564 (1971).
14. N.M. Concessio and A.J. Hickey, "Analysis of Patterns of Particle Dispersion from a Dry Powder Inhaler," *Pharm. Technol.* 20 (6), 58-62 (1996).
15. Elsons Corporation, "Intal Cromolyn Sodium: A Monograph," 1973.
16. N.M. Concessio et al., "Factors Influencing the Dispersion of Dry Powders as Aerosols," *Pharm. Technol.* 18 (9), 58-82 (1994).
17. N.M. Kassem, "Generation of Deeply Inspirable Clouds from Dry Powder Mixtures," PhD thesis, University of London, 1990.
18. K.C. Lee et al., "Investigation of the Aerodynamic Characteristics of Inhaler Aerosols with an Inhalation Simulation Machine," *Int. J. Pharm.* 130, 103-113 (1996).
19. M.P. Timmins et al., "Drug Delivery to the Respiratory Tract Using Dry Powder Inhalers," *Int. J. Pharm.* 101, 1-13 (1994).
20. M.J. Kontny, J.J. Connors, and E.T. Graham, "Moisture Distribution and Packaging of Dry Powder Systems," in *Respiratory Drug Delivery V*, P.R. Byron, R.N. Dalby, and S.J. Farr, Eds. (Interpharm Press, Buffalo Grove, 1996), pp. 125-136.
21. A.R. Clark and R.B. Bailey, "Inspiratory Flow Profiles in Disease and Their Effects on the Delivery Characteristics of Dry Powder Inhalers," in *Respiratory Drug Delivery V*, P.R. Byron, R.N. Dalby, and S.J. Farr, Eds. (Interpharm Press, Buffalo Grove, 1996), pp. 221-230.
22. A.J. Hickey, "Lung Deposition and Clearance of Pharmaceutical Aerosols: What Can Be Learned from Inhalation Toxicology and Industrial Hygiene," *Aerosol Sci. Technol.* 18, 290-304 (1993).
23. M. Sacchetti and M.M. Van Cort, "Spray Drying and Supercritical Fluid Particle Generation Techniques," in *Inhalation Aerosols: Physical and Biological Basis for Therapy*, A.J. Hickey, Ed., Marcel Dekker, N.Y., 1996, pp. 337-384.
24. J.W. Tom and P.G. Debenedetti, "Particle Formation with Supercritical Fluids — A Review," *J. Aerosol Sci.* 22, 555-584 (1991).
25. E.M. Phillips and V.J. Stella, "Rapid Expansion from Supercritical Solutions: Application to Pharmaceutical Processes," *Int. J. Pharm.* 94, 1-10 (1993).
26. P. York and M. Hanna, "Particle Engineering by Supercritical Fluid Technologies for Powder Inhalation Drug Delivery," in *Respiratory Drug Delivery V*, P.R. Byron, R.N. Dalby, and S.J. Farr, Eds. (Interpharm Press, Buffalo Grove, 1996), pp. 231-239.
27. R. Falk et al., "Controlled Release of Ionic Compounds from Poly (L-lactide) Microspheres Produced by Precipitation with a Compressed Antisolvent," *J. Controlled Release* 44, 77-85 (1997).
28. T.W. Randolph et al., "Sub-Micrometer-Sized Biodegradable Particles of Poly (L-lactic Acid) via the Gas Antisolvent Spray Precipitation Process," *Bio-technol. Prog.* 9, 429-435 (1993).
29. R. Bodmeier et al., "Polymeric Microspheres Prepared by Spraying into Compressed Carbon Dioxide," *Pharm. Res.* 13 (8), 1211-1217 (1996).
30. B. Zierenberg et al., "Boehringer Ingelheim Nebulizer BIneb: A New Approach to Inhalation Therapy," in *Respiratory Drug Delivery V*, P.R. Byron, R.N. Dalby, and S.J. Farr, Eds. (Interpharm Press, Buffalo Grove, 1996), pp. 187-193.
31. S.J. Farr et al., "ABRx: Development of a Novel Liquid Aerosol Delivery System: Concept to Clinic," in *Respiratory Drug Delivery V*, P.R. Byron, R.N. Dalby, and S.J. Farr, Eds. (Interpharm Press, Buffalo Grove, 1996), pp. 175-185.
32. P. Lloyd et al., "A New Unit Dose, Breath-Actuated Aerosol Drug Delivery System," in *Respiratory Drug Delivery V*, P.R. Byron, R.N. Dalby, and S.J. Farr, Eds. (Interpharm Press, Buffalo Grove, 1996), pp. 364-366.
33. P.J. Atkins, "Aerodynamic Particle-Size Testing — Impinger Methods," *Pharm. Technol.* 16 (8), 26-32 (1992).
34. S.M. Milosovich, "Particle-Size Determination via Cascade Impaction," *Pharm. Technol.* 16 (9), 82-86 (1992).
35. J. Ranucci, "Dynamic Plume-Particle Size Analysis Using Laser Diffraction," *Pharm. Technol.* 16 (10), 109-114 (1992).
36. R.W. Niven, "Aerodynamic Particle Size Testing Using a Flow-of-Flight Aerosol Beam Spectrometer," *Pharm. Technol.* 17 (1), 72-78 (1993).
37. W.G. Gorman and F.A. Carroll, "Aerosol Particle-Size Determination Using Holography," *Pharm. Technol.* 17 (2), 34-37 (1993).
38. R. Evans, "Determination of Drug Particle Size and Morphology Using Optical Microscopy," *Pharm. Technol.* 17 (3), 146-151 (1993).
39. P.D. Jager, G.A. Da Stefano, and D.P. McNamara, "Particle-Size Measurement Using Right-Angle Light Scattering," *Pharm. Technol.* 17 (4), 102-120 (1993).
40. J.A. Ranucci and F.C. Chen, "Phase Doppler Anemometry: A Technique for Determining Aerosol Plume-Particle Size and Velocity," *Pharm. Technol.* 17 (6), 62-74 (1993).
41. Task Group on Lung Dynamics, "Deposition and Retention Models for Internal Dosimetry of the Human Respiratory Tract," *Health Physics* 12, 173-208 (1966).
42. I. Gonda, "Semi-Empirical Model of Aerosol Deposition in the Human Respiratory Tract for Mouth Inhalation," *J. Pharm. Pharmacol.* 33, 692-696 (1981).
43. G.W. Hallworth and D.G. Westmoreland, "The Twin Impinger: A Simple Device for Assessing the Delivery of Drugs from Metered-Dose Pressurized Aerosol Inhalers," *J. Pharm. Pharmacol.* 38, 966-972 (1987).
44. R. Jager-Waldau, "The Use of the Twin Impinger in Evaluating Pharmaceutical Aerosols," unpublished data, 1995.
45. S.J. Dunnett and D.B. Ingham, "The Human Head as a Breathing Aerosol Sampler," *J. Aerosol Sci.* 19 (3), 365-380 (1988).
46. V.A. Marple, "Simulation of Respirable Penetration Characteristics,"

Sacchetti and M.M. Van Oort, "Spray Drying and Supercritical and Particle Generation Techniques," in *Inhalation Aerosols: Physical and Biological Basis for Therapy*, A.J. Hickey, Ed., Oxford, NY, 1996, pp. 337-384.

V. Tom and P.G. Debenediti, "Particle Formation with Supercritical Fluids—A Review," *J. Aerosol Sci.* 22, 555-584 (1991).

V. Phillips and V.J. Stella, "Rapid Expansion from Supercritical Solutions: Application to Pharmaceutical Processes," *Int. J. Pharm.* 94, 1-10 (1993).

York and M. Hanne, "Particle Engineering by Supercritical Fluid Technologies for Powder Inhalation Drug Delivery," in *Respiratory Drug Delivery*, V.P.R. Byron, R.N. Dalby, and S.J. Farr, Eds., Interpharm Press, Buffalo Grove, 1996, pp. 231-239.

Falk et al., "Controlled Release of Ionic Compounds from Poly(lactide) Microspheres Produced by Precipitation with a Compressed Antisolvent," *J. Controlled Release* 44, 77-83 (1997).

J. Randolph et al., "Sub-Micrometer-Sized Biodegradable Particles of Poly(L-Lactic Acid) via the Gas Antisolvent Spray Precipitation Process," *Biotechnol. Prog.* 9, 429-435 (1993).

Kodmeyer et al., "Polymeric Microspheres Prepared by Spraying Compressed Carbon Dioxide," *J. Pharm. Res.* 13 (8), 1211-1217 (1996).

Sierenberg et al., "Boehringer Ingelheim Nebulizer BIneb: A New Approach to Inhalation Therapy," in *Respiratory Drug Delivery*, V.P.R. Byron, R.N. Dalby, and S.J. Farr, Eds., Interpharm Press, Buffalo Grove, 1996, pp. 187-193.

Farr et al., "AERx: Development of a Novel Liquid Aerosol Delivery System: Concept to Clinic," in *Respiratory Drug Delivery*, V.P.R. Byron, R.N. Dalby, and S.J. Farr, Eds., Interpharm Press, Buffalo Grove, 1996, pp. 175-185.

Lloyd et al., "A New Unit Dose, Breath-Actuated Aerosol Drug Delivery System," in *Respiratory Drug Delivery*, V.P.R. Byron, R.N. Dalby, and S.J. Farr, Eds., Interpharm Press, Buffalo Grove, 1996, pp. 364-366.

Atkins, "Aerodynamic Particle-Size Testing—Impinger Method," *Pharm. Technol.* 16 (8), 26-32 (1992).

Milosevich, "Particle-Size Determination via Cascade Impaction," *Pharm. Technol.* 16 (9), 52-56 (1992).

Inucci, "Dynamic Plume Particle Size Analysis Using Laser Interference," *Pharm. Technol.* 16 (10), 109-114 (1992).

Niven, "Aerodynamic Particle Size Testing Using a Time-Flight Aerosol Beam Spectrometer," *Pharm. Technol.* 17 (1), 8 (1993).

Gorman and F.A. Carroll, "Aerosol Particle-Size Determination Using Holography," *Pharm. Technol.* 17 (2), 34-37 (1993).

Wang, "Determination of Drug Particle Size and Morphology by Optical Microscopy," *Pharm. Technol.* 17 (3), 146-152 (1993).

Jager, G.A. De Stefano, and D.P. McNamara, "Particle-Size Measurement Using Right-Angle Light Scattering," *Pharm. Technol.* 17 (4), 102-120 (1993).

Ranucci and F.-C. Chen, "Phase Doppler Anemometry: A Technique for Determining Aerosol Plume-Particle Size and Velocity," *Pharm. Technol.* 17 (6), 62-74 (1993).

Group on Lung Dynamics, "Deposition and Retention Model: Internal Dosimetry of the Human Respiratory Tract," *Health Res.* 12, 173-208 (1966).

Wada, "Semi-Empirical Model of Aerosol Deposition in the Human Respiratory Tract for Mouth Inhalation," *J. Pharm. Pharmacol.* 33, 692-696 (1981).

Hallworth and D.G. Westmoreland, "The Twin Impinger: A New Device for Assessing the Delivery of Drugs from Metered Pressurized Aerosol Inhalers," *J. Pharm. Pharmacol.* 39, 172 (1987).

Ger-Waldau, "The Use of the Twin Impinger in Evaluating Inhalation Aerosols," unpublished data, 1993.

Jannett and D.B. Ingham, "The Human Head as a Blunt or Sampler," *J. Aerosol Sci.* 19 (3), 365-380 (1988).

Marple, "Simulation of Respirable Penetration Characteristics

- by Inertial Impaction," *J. Aerosol Sci.* 9, 125-134 (1978).
- J.P. Newman et al., "Deposition of Pressurized Aerosols in the Human Respiratory Tract," *Thromb.* 36, 52-55 (1981).
- G. Rudolf, R. Kobrich, and W. Stahlhofen, "Modeling and Algebraic Formulation of Regional Aerosol Deposition in Man," *J. Aerosol Sci.* 21, 5403-406 (1990).
- USP 22 Aerosols (601), Seventh Supplement (The United States Pharmacopoeial Convention, Rockville, MD, 1992), pp. 2123-3129.
- M. Van Oort, "In Vitro Testing of Dry Powder Inhalers," *Aerosol Sci. Technol.* 22, 364-373 (1995).
- V.A. Marple, B.A. Olson, and N.C. Miller, "A Low-Loss Cascade Impactor with Stage Collection Cups: Calibration and Pharmaceutical Inhaler Applications," *Aerosol Sci. Technol.* 22, 124-134 (1995).
- R.N. Dalby and S.L. Tiano, "Pitfalls and Opportunities in the Inertial Sizing and Output Testing of Nebulizers," *Pharm. Technol.* 17 (9), 144-156 (1993).
- A.R. Clark, "Metered Atomization for Respiratory Drug Delivery," PhD thesis, Loughborough Univ. Tech., Loughborough, UK, 1991.
- C.A. Dunbar, "An Experimental and Theoretical Investigation of the Spray Issued from a Pressurized Metered-Dose Inhaler," PhD thesis, UMIST, Manchester, UK, 1996.
- D.S. Junc, N.C. Miller, and R.K. Schultz, "A Conceptual Model for the Development of Pressurized Metered-Dose Hydrofluoroalkane-Based Inhalation Aerosols," *Pharm. Technol.* 18 (10), 40-52 (1994).
- M. Hyland, "A Spoonful of Sugar?—Asthma Drug Delivery and Compliance, Improving Drug Delivery from Pressurized Inhalers," International Pharma Technology Programme, London, 1995.
- P.R. Byron and J.S. Patton, "Drug Delivery via the Respiratory Tract," *J. Aerosol Med.* 7 (1), 49-75 (1994).
- R.W. Niven, "Delivery of Biopharmaceuticals by Inhalation Aerosols," *Pharm. Technol.* 17, 72-82 (1993).

NEWS BRIEF

Pump efficiency video available from DOE

The US Department of Energy (DOE) and the Hydraulic Institute (HI), a trade association of pump manufacturers, have created an educational video to promote energy efficiency in electrically powered pumping systems.

According to DOE, pumps and pumping systems account for more than 5% of the total energy used in the United States. The savings potential based on educating and training pump users and engineering contractors in the selection and use of pumps is estimated at 20%.

The video identifies specific opportunities for energy savings and explains the calculations and adjustments necessary to achieve higher efficiency. The hour-long program provides information regarding pump systems design, evaluation of performance characteristics, avoidance of excessive capacity and total head margins, and the use of multiple pumps and recovery turbines.

Production of the video was funded by DOE's Motor Challenge Program, an industry and government partnership designed to help industry save 5 billion kWh of electricity annually by 2000. The Motor Challenge promotes industrial energy efficiency through the use of efficient electric motors, drives, and drive equipment, and effective electric motor-driven system integration and optimization.

For further information on the DOE/HI video program, the Motor Challenge program, or to obtain the pump efficiency video and materials, contact the Motor Challenge Clearinghouse at 1301 Clay St., Oakland, CA 94612, tel. (800) 862-2056.

Biofil Polyethersulphone Membrane Cartridge Filters




NAFCO Biofil (single and multiple cartridges) are engineered for high strength, long life and precise particle retention. Biofil cartridges are manufactured in certified clean rooms to the highest ISO8001 quality assurance standards. All materials conform to FDA requirements for bio-safety.

Call NAFCO today for a comprehensive validation guide of Biofil cartridges! Phone: 315-331-9100

FAX: 315-331-3730 Circle 72

North American Filter Corporation • 200 Westshore Blvd. • Newark, New York 14513

Do you require particle size results of *highest resolution* ... like this?



If so, you need the right kind of particle size analyzer - one that doesn't miss the details. Our NEW Acoustar™ 770 is different from all the rest. Designed with proven optical technology, innovative electronics and Autofocus™, the 770 succeeds where alternate technologies fail.

- Wide dynamic size range: 0.5-2600 microns
- Autofocus (pat.), ashram-compatible
- Fast cycle time (< 5 min.), easy to use
- Automatic ONLINE sampling capability
- Compatible with NICOPT™ 370 Submicron Particle Sizer module, 0.003 - 5 µm
- NEW - Zeta Potential modular option
- NEW - Dry Powder Feed Option

PSS PARTICLE SIZING SYSTEMS
75 Aero Camino • Santa Barbara, CA 93117
Tel: (805) 968-1497 FAX: (805) 968-0361

Circle 12